Operation and Maintenance Manual

Flowmaster 250 MK2



Contents


1	GETTING TO KNOW YOUR FLOWMASTER			
	1.1	What's in the box?		
2	SPECIFICATION			
	2.1	General		
	2.2 2.3	Flowmeter Display		
	2.4	Battery Pack		
	2.5	Software		
	2.6	Further Manuals		
3	FLOWMASTER OPERATION			
	3.1	Flowmeter Display Power On Sequence		
	3.2	Explanation of Display Panel		
4	USING YOUR FLOWMASTER			
	4.1 4.2	Simple Steps to Fast Flow Analysis Precautions for Use		
5	MA	8		
	5.1	Battery Maintenance		
	5.2 5.3	Flowmaster Maintenance Troubleshooting		
	5.4	Calibration & Resetting Calibration Due Date Reminder Message		
6	LI ION BATTERY CHARGER			
	6.1 How to Use the Charger 6.2 FAQ & Trouble-shooting			
7	ACCESSORIES			
8	LIMITED WARRANTY			
9	CE DECLARATION OF CONFORMITY			
10	CA	LIBRATION CERTIFICATE	13	
10	CA	LIBRATION CERTIFICATE	13	

I Getting to Know Your Flowmaster



Battery Charger





2 Specification

2.1 GENERAL

Weight 12.5 kg including battery

pack & instantaneous adaptors Recharging

Dimensions 342 (L) x 146 (W)

x 336 (H) mm

Power Internal Battery Pack. DC 12V @

250 mA

Case Material Aluminium
Operating Temp. -10°C ~ +50°C

Adaptors 2.5" BSPM or BS Instantaneous

Storz B, Storz 65, 2.5" NST adaptors available on request

2.2 FLOWMETER

Colour Firemen Red & anodised

aluminium

Sensor Type Electromagnetic Range 30 - 3,000 lpm

By changing the sensor control pcb, range can be extended to

5,000 lpm

Resolution 50 ~ 750 l/min ± 15 l/min

750 l/min ~ ±2%

Standard Functions Flow reading

Totaliser reading

Casing Material Corrosion resistant ADC6-12 (JIS)

aluminium (painted)

Pressure rating 16 bar

2.4 BATTERY PACK

Battery Type Lithium-Ion, Li Ion

Rated Capacity 2.4 Ahr (10 hours operation)
Recharging Mains operated recharger is

supplied

In-car charger Available as separate item

Monitor A built-in battery monitor displays

remaining capacity in hrs/mins

External Battery A

Pack

Available as separate item

2.5 MANUALS

The following manuals are provided on a CD supplied with the Flowmaster:

- · Operation and Maintenance Manual
- You Tube videos demonstrating how to use the Flowmaster 250 MK2
- · Quick start guide



2.3 DISPLAY

Display Type 320 x 240 backlit graphic

colour display

All readings are available on this

display.

Flow sensor There are four flow unit selections

to choose from; - LPM - m3/h - LPS - US GPM

Pressure Transducer Bar, PSI, KPa, MTrH readouts

selectable

Type - Strain Gauge Range - 0 ~ 25 bar, ±1%

Volume Totaliser Showcase of total volume.

Sortable by test and group, as well as keeping track of lifetime

total

Calibration Sensor calibration is accessible

via the password protected

engineering page


3 Flowmaster Operation

3.1 DISPLAY PANEL POWER-ON SEQUENCE

Turn on the Flowmaster by pressing the ON/OFF switch to "I" position. When the Flowmaster is first turned on, the flow meter's self-test mode will commence. This can last up to 5 seconds, after which the Flowmaster will be booted and ready for use.

3.2 EXPLANATION OF DISPLAY PANEL

The display panel features a 320 x 240 pixel LCD panel. Beneath the LCD panel, there are four buttons (icons above the buttons indicate their function), a joystick and an OK button, as shown below.



Parameter Settings Menu

Flow Units Selection Menu

Press button 3 on home page to access. Navigate using the joystick. Use button 1 to return to home page, use button 4 to proceed to next menu.



Pressure and Volume Selection Menu

Use the up-down arrows on the joystick to change pressure and volume units. You can navigate the options by the left-right joysticks.



Battery Info Menu

In this menu, you can view information about battery life and details about the charges, voltage and internal battery level.



4 Using Your Flowmaster

4.1 SIMPLE STEPS TO FAST FLOW ANALYSIS



Measure and record pressure and flow rate



IMPORTANT - PLEASE READ THIS SECTION BEFORE USING YOUR FLOWMASTER.

It is important to think safety first, especially when working with pressure.

- Please ensure the Flowmaster is securely attached to the service fitting (eg. hydrant outlet, pump, etc.) before pressurising.
- Please consider carefully the expected system pressures if you plan to use a length of hose between the service point and the Flowmaster.
- **Never** allow the Flowmaster to be used in an open-ended or non-fixed position. It is essential to secure the Flowmaster so that it cannot move about freely, as this may result in injury and/or Flowmaster damage.

4.5 VIDEO TUTORIALS



Setting Up on a Hydrant



Setting Up In-Line Between Two Hose Lengths



Understanding Readouts While There Are Readings

5 Maintenance & Troubleshooting

5.1 BATTERY MAINTENANCE

- The battery pack operating time is 10 hours continuous usage. It is better to top up charge the batteries at regular intervals than to allow full discharge.
- The battery type is designed for regular but intermittent use. It is not ready for deep discharge that would come from leaving the Flowmaster powered on for eight to ten hours a day. If your procedures require this activity, please contact TSI Flowmeters and request alternative battery technology.
- The flow measurement will become erratic if the battery voltage is too low (readout indicating zero when water is flowing). It's recommended to use a voltmeter to check the voltage of the battery pack. If less than 10V DC, immediate recharging is necessary. If less than 12V DC, immediate top-up charging is recommended.
- Batteries have an operating life of 10 years and will need to be replaced thereafter.

5.2 FLOWMASTER MAINTENANCE

- Avoid dropping the Flowmaster. Use your hands to operate the switches and avoid use of screwdrivers. Keep abrasive chemicals away from the exterior. Wash the inner tube with clean water after.
- The flow meter control electronics are in a waterproof compartment located beneath the display gauge Do'not open this compartment. This will destroy the waterproof feature and void warranty.
- The Flowmaster's flow gauge response is reduced if the electrodes are covered with a film of oil, dirt or rust. This can result in measurement error. It is recommended to clean the sensor electrodes periodically - Clean the electrodes by brushing the inside of the pipe with a soft cloth or soft brush dipped in water. If river water is used, it's recommended to clean the electrodes every 6 months. If oily water has been used, washed immediately.
- The Flowmaster is calibrated and scale adjusted at the factory, therefore accurate measurements can be made immediately upon receipt of the product. It is recommended, however, that all measuring instruments are calibrated annually.

5.3 TROUBLESHOOTING

Incomplete Self-Test Sequence: The LED displays a series of dashes after self-test sequence commences.

Possible Cause: The Sensor and Display parts of the device are not in contact.

Action: Check all cables are secure. If not connected, contact your local representative.

Erratic LED Display

Possible Cause: The battery voltage is too low.

Action: Charge the batteries.

Reading Not Displaying: The self-test sequence is okay, but there is no reading when water is flowing. **Possible Cause:** 1. Batteries voltage is too low. 2. No earth wire to battery. 3. Unit is faulty.

3. Return to manfacturer. **Action:** I. Charge or replace batteries. 2. Replace earth wire between

negative terminal and red casting.

Recharging Issues: The unit will not charge or there are humming noises while trying. **Possible Cause:** Faulty battery charger ot batteries are over-discharged and cannot recover.

Action: Contact local representative for replacement charger or batteries.

All log file dates are Year 2000: The data logger's de-charged battery will revert to default date of the year 2000.

Possible Cause: Battery not charged.

Action: Leave Flowmaster powered on for at least 3 hours, then set the date and time.


If the above action does not work, return your meter to TSI for repair.

5.4 CALIBRATION & RESETTING CALIBRATION DUE DATE REMINDER MESSAGE

Every TSI flow meter is tested and calibrated prior to shipping. A calibration label is attached to the top of the

meter, and on this label will be the date on which the meter is due its next calibration.

The master meter at TSI is calibrated annually at the Irish National Test Centre, this ensures the reading obtained by the Test Equipment referred to on certificates issued by TSI Flowmeters Ltd are traceable to National and International standards.



6 Li Ion Battery Charger

6.1 How to use the charger

* It is recommended to power off the Flowmaster while charging the Li Ion battery. *

- The charger is started by connecting the battery packer to the charger and then connecting the charger to the mains.
- When the mains is connected the LED will be orange for approximately 8 seconds before turning to yellow when initialisation and analysis starts.
- The LED turning to orange indicates the beginning of Fast Charge. Once the batteries are fully charged and the voltage drops, the charger will go into a Top-off Charge mode.
- When the LED turns a constant green, the battery has entered Trickle Charge. It is recommended that trickle charge does not exceed 24 hours.

LED	MODE	
YELLOW	Battery not connected	
YELLOW	Battery initialisation and analysis	
ORANGE	Fast charge	
GREEN with intermittent YELLOW flash	Top-Off Charge	
GREEN	Trickle Charge	
Alternating ORANGE-GREEN	ERROR	

If the mains input voltage is turned off, the charger will reset. When the voltage is turned on again, a new charge cycle will start.

If new batteries are to be connected, the charger must idle for approx. I5 seconds to make sure all parameters in the microprocessor have been reset. When the charger has been reset the LED changes to yellow, and a new charge cycle can begin.

The Li Ion Battery



6.2 FAQ and Trouble-shooting

Battery does not seem to charge

Making a good electrical connection between charger plug and battery socket: Observe the LED indication to confirm the electrical connection between charger and battery. Redo the insertion of charger plug into meter socket, power off and on the charger and observe until a 'normal' charge cycle is achieved.

Battery charges very quickly or meter stops working after a short period of time

A full charge cycle from empty takes ~2.5 hours. If the charging cycle (from empty) is significantly briefer, it may indicate that the battery is nearing the end of its life. In this case, battery replacement is the recommendation.

Memory effect

As far as TSI is aware, there is NO memory effect with Li Ion batteries. It is acceptable to charge the batteries from any state of capacity.

How many charge cycles over life time of an Li Ion battery

Battery manufacturer literature suggests that after between 300 and 500 discharge and recharge cycles, the remaining battery capacity will be ~75% of initial rated capacity.

7 Accessories

7.1 BALL VALVE FITTINGS

The ball valve facilitates speedy and safe static pressure readings and creates back pressure for taking flow and residual pressure measurements









BS Inst.

Storz

7.2 TRANSPORT & STORAGE CASE

Transport case with castor wheels and extension handle. Foam interior cut to hold Flowmaster and accessories. Weight - 12.5kg.



7.3 ADAPTORS

A wide range of adaptors are available that allow the Flowmaster to work with your pipework and fittings.

NH

BS Instantaneous

Flange to 2.5" BSP Male

Flage to 4" BSP Male Storz 65

Storz A

Storz B















Storz C

Guillemin 100

Guillemin 65

Barcelona

NOR Kuppling

GOST













7.4 EXTERNAL BATTERY/CHARGERS



External Li Ion Battery Pack available



Mains Battery Charger available



In-car Charger available

8 Limited Warranty

TSI Flowmeters Ltd., of Unit 1, Portlaoise Enterprise Centre, Clonminam Business Park, Portlaoise, Co. Laois, R32VK07, Ireland (Warrantor), warrants to the original purchaser of the new fire protection equipment manufactured by Warrantor and to any person to whom such equipment is transferred, that such equipment shall be free from defects in materials and workmanship during the one (1) year period commencing upon the receipt of such equipment by the original purchaser thereof ("warranty period").

Warrantor's obligation under this warranty is specifically limited to replacing or repairing its fire protection equipment or parts thereof, which are shown by Warrantor's examination to be in a defective condition attributable hereunder to Warrantor. To qualify for this Warranty, alleged defective equipment MUST be returned to Warrantor at its above address, transportation charges prepaid, within a reasonable time after discovery of an alleged defect, and in no event later than thirty (30) days after the expiration of the warranty period. If, as a result of Warrantor's examination of returned equipment, Warrantor concludes that a product defect attributable hereunder to Warrantor exists, Warrantor shall cure such defect within a reasonable time, not to exceed forty-five (45) days after such examination. All expenses in curing such defect, except for transportation charges and shipping expenses incurred in delivering such equipment to Warrantor, shall be paid by Warrantor.

In the event that such equipment is found to be attributable hereunder to Warrantor and Warrantor is unable to provide replacement, or repair is not commercially practicable or cannot be timely made, Warrantor may elect to refund to claimant the purchase price of such equipment actually received by Warrantor, less reasonable depreciation, in complete discharge of its obligations hereunder. If Warrantor elects to comply with this warranty by means of such refund, as a condition precedent to such compliance, the claimant shall return such equipment to Warrantor free and clear of liens and other encumbrances.

THE ORIGINAL PURCHASER OF SUCH EQUIPMENT, AND PERSON TO WHOM SUCH EQUIPMENT IS TRANSFERRED, AND ANY PERSON WHO IS AN INTENDED OR UNINTENDED BENEFICIARY OF SUCH EQUIPMENT, SHALL NOT BE ENTITLED TO RECOVER FROM WARRANTOR ANY CONSEQUENTIAL OR INCIDENTAL DAMAGES FOR INJURY TO PERSON AND/OR PROPERTY RESULTING FROM ANY DEFECTIVE EQUIPMENT MANUFACTURED BY WARRANTOR.

Misuse or neglect (including failure to provide reasonable maintenance) of, or accident or unauthorised repairs or alterations to, such equipment, shall release and discharge Warrantor from any obligations under this warranty or otherwise.

WARRANTOR EXPRESSLY LIMITS WITH RESPECT TO SUCH EQUIPMENT ALL IMPLIED WARRANTIES OF MERCHANTABILITY AND ALL IMPLIED WARRANTIES OF FITNESS FOR A PARTICULAR PURPOSE TO THE WARRANTY PERIOD. AFTER EXPIRATION OF THE WARRANTY PERIOD, WARRANTOR EXPRESSLY DISCLAIMS WITH RESPECT TO SUCH EQUIPMENT ALL IMPLIED WARRANTIES OF MERCHANTABILITY AND ALL IMPLIED WARRANTIES OF FITNESS FOR A PARTICULAR PURPOSE. THERE IS NO WARRANTY OF ANY NATURE MADE BY WARRANTOR BEYOND THAT WHICH IS CONTAINED HEREIN.

Should Warrantor fail to meet with its obligations under this warranty, a claimant may sue Warrantor to secure its compliance with this warranty. No action to enforce this warranty or to otherwise secure recovery from Warrantor for any damages arising out of the fire protection equipment manufactured by Warrantor shall be commenced later than two (2) months from and after the date of the receipt of such equipment by the original purchaser thereof.

NO PERSON HAS AUTHORITY TO ENLARGE, AMEND, OR MODIFY THIS WARRANTY.

Warrantor reserves the right to change the parts or design of its products from time to time without notice, and with no obligation to maintain spare parts or to make corresponding changes in the products previously manufactured.

9 CE Declaration of Conformity



We: TSI Flowmeters

Unit 1, Portlaoise Enterprise Centre, Clonminam Business Park, Portlaoise, Co. Laois, R32 VK07, Ireland.

Tel +353 (0)57 866 3852

declare under our sole responsibility that the product,

Flowmaster 250 MK2

to which this documentation relates, is in conformity with the provisions of the following directives:

2001/95/EC General Product Safety 2006/95/EC Low Voltage Directive (LVD) 2004/108/EC Electromagnetic Compatibility Directive (EMC)

The technical file is maintained at:

TSI Flowmeters Ltd, Unit 1, Portlaoise Enterprise Centre, Clonminam Business Park, Portlaoise, Co. Laois, R32 VK07, Ireland.

Date of Issue: 10th August 2019

Place of Issue: Portlaoise, Ireland

Tim Corw.

Name: **Tim Carew**

Title: Managing Director

10 Calibration Certificate

Issued by:		_					
TSI Flowmete		Unit 1, Portlaoise Enterprise Centre, Clonminam Business Park, Portlaoise, Co. Laois, R32 VK07, Ireland. Tel: +353 (0)57 866 3852 Email: info@tsi.ie					
The readings obtained by the Test Equipment referred to on this certificate are traceable to National and International standards. Calibration Cert No. 6825CN.							
Cert No.	Date of Issue	Order No.	Job No.				
-01							
Company Name:			- - -				
Details of Flowmeter Calibrated							
Type of Meter: Serial No. Date of Calibration:	TSI Model EMF-300E	Magflow					
Method of Test: Comparison with Master Flowmeter.							
The zero offset was adjusted to read 0 lpm. Since this is the first calibration of a new flow meter there are no before service readings. The unit was calibrated at a flow rate of 500 lpm. This flow rate was maintained for five minutes and correct reading was observed on the unit. Totaliser readings were taken at one minute intervals; the maximum deviation observed was within the specified $\pm 2\%$ of master flow meter readings. Correct flow rate was verified for this calibration setting at 300, 400, 600 and 700 lpm.							
Calibrated by:	TSI Flowmeters Ltd						

Certificate of Calibration

Operation & Manual Maintenance Manual

Flowmaster 250 MK2

Portable Flow, Volume and Pressure measuring instrument